## BMT202 MESLEKI YABANCI DIL 2 11.HAFTA

GAMA MESLEK YÜKSEKOKULU ÖĞR.GÖR.MEHMET DURSUN

Performance Verification
Introduction
This section discusses the tests used to verify performance following repairs or during routine
maintenance.

All tests can be performed without removing the covers of the monitor.
If the monitor fails to perform as specified in any test, repairs must correct the problem before
the monitor is returned to the user.
Performance Tests

Battery Charge
To fully charge the battery:
1. Connect the monitor to an AC power source using the proper power cord.
For the A1, use the PS-120V or PS-240V external power supply and power cord.
1. Connect the monitor to an AC power source using the proper power cord.

2. For the A1, verify that the EXTERNAL POWER indicator is lit.
For the A3, verify that the BATTERY CHARGING/AC SOURCE indicator is lit.
3. Charge the battery for at least 8 hours.
The battery can require a complete discharge/charge cycle to restore its normal capacity,
depending on its previous usage.

1. To check for a full charge, perform the procedure described in Battery Performance T	est or
page 12.	

Kaynakça:

Philips A1 and A3 Patient Monitors- Service Guide

Service Guide.

**Philips Medical Systems** 

https://www.usa.philips.com/healthcare